

**बिड दस्तावेज़ / Bid Document**

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	13-12-2025 17:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	13-12-2025 17:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	30 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	Gujarat
विभाग का नाम / Department Name	Panchayats And Rural Housing Department Gujarat
संगठन का नाम / Organisation Name	Gandhinagar District Panchayat
कार्यालय का नाम / Office Name	Health
कुल मात्रा / Total Quantity	15
वस्तु श्रेणी / Item Category	Deep Freezer For Universal Immunization Program (Q3)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	3 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) / OEM Average Turnover (Last 3 Years)	3 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service	3 Year (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / MSE Exemption for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Exemption for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़ / Document required from seller	Experience Criteria, Past Performance, Bidder Turnover, OEM Authorization Certificate, OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

बिड विवरण/Bid Details	
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	10 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य /Estimated Bid Value	698500
Payment Timelines	Payments shall be made to the Seller within <b>45</b> days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

#### ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	13440

#### ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) / ePBG Percentage (%)	0.01
ईपीबीजी की आवश्यक अवधि (माह) / Duration of ePBG required (Months).	6

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

#### लाभार्थी /Beneficiary :

AO

Health, Panchayats and Rural Housing Department Gujarat, Gandhinagar District Panchayat,  
(Shantaben Rameshbhai Garasiya)

#### UIN Number NCTGC2415P

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--------------------------------------------	----

#### एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	25

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as

indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25 % percentage of total quantity. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023. [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

6. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

7. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

## Deep Freezer For Universal Immunization Program ( 15 pieces )

### तकनीकी विशिष्टियाँ /Technical Specifications

[\\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Deep Freezer Technical Parameter	Refrigerant Type	
	Inlet Of Capillary Should Be Outside The PUF Body	
	ON/OFF Switch And Power Indicator Should Be Available	

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Type	
	It Should Have Foam Pad Cover On Top Of The Basket (Applicable For Chest Type)	
	Gross Volume	(liter)
	Solid Door With Lock And Handle	
Deep Freezer Construction	External Material Corrosion Resistance	Yes
Deep Freezer Operational Requirements	Ground Clearance Of Unit	100.0 (millimeter)
	Manufacturing process of the product should not use or produce hazardous chemicals-gases.	Yes
Deep freezer Technical Parameter	A Vertical Arrow Should be Marked At The All Sides Of Packages To Ensure Transportation Of Equipment In Vertical Position, TOP and BOTTOM Should Also Be Written	Yes
	Power Input	220-240 VAC; 50-60Hz
	<b>Testing should be carried out from WHO certified lab/NABL//STQC Lab</b>	Yes
	<b>Standards and Safety</b>	WHO, UNICEF Standard, PQS, E03, FZ03.1
	Micro-Processor System Configuration Features	Programmable Micro-processor Control Unit With Child Lock Facility, Should have provision to set minimum and maximum temperature at 0.1 degree Centigrade to programme the unit for continuous operation, Should have provision for defrosting program
	Micro processor based control panel features	3 digit digital display, Sensor be placed 25 to 50mm above the chamber, Power on LED/LCD indicator, Audio (minimum 65 dBA) and visual alarm against the violation of set temperature range, Temperature manual control with one decimal degree scale, The unit should be sealed protected from dust, moisture or condensed water falling over it, Sensor Accuracy for digital controller +/- 0.5 degree centigrade

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>Warranty, Documentation and Packaging</b>	The Equipment Should have Minimum Warrantee Including Comprehensive Maintenance Of Sixty months after installation or Sixty Six Months after the supply whichever is later, A) A paper copy of User/Operator Manuals to be supplied in English. B)A paper copy of technical/wiring diagram/maintenance manuals to be supplied in English. C) Certificate of Inspection for technical compliance from an independent laboratory approved/recognized by WHO certified/NABL/STQC lab is essential. Certificate of testing should be currently valid till the supply and same must be verified by inspecting authority. D) List of important spare parts and accessories with their part number and costing., A) The Supplier should provide strong and sufficient packing to ensure safe arrival of goods at the destination free from loss or damage. B) To put label and signage's for HANDLE WITH CARE ON ALL SIDES OF THE CRATES as per packing an shipment norms., The Supplier Required to maintain all the spare parts throughout the warrantee period and not less than 10 Years
Alcohol Stem Thermometer Technical Parameter	<b>Lower Limit Temperature Ranges</b>	-30 deg Celcius
	Construction Of alcohol Stem Thermometer	The Glass Column Must Be Protected Against Breakage and strongly supported, The column cannot be displaced more than 0.5mm vertically with respect to the scale, The reading angle is between 80/100degree to plane of support plate
	<b>Upper Limit Temperature Ranges</b>	+50 deg Celcius
	<b>Vibration test</b>	Product should stand 30 minutes on a programmable vibrating table without physical damage orcalibration, Calibration certificate from NABL/STQC/WHO approved labshould be submitted
	Casing	Non-Corrodible
	Mounting	Hook to Suspend From Shelf Or Adhesive, Used inside refrigerators or freezers in small health centres, Can also be packed with vaccines during transport, WHO Specification reference: E6/TH.3, Applies to procedures:E6/PROC/2
	<b>Temperature Accuracy</b>	+/- 1 deg Celcius
	<b>Warranty of Alcohol Stem Thermometer</b>	The warranty Shall Remain Valid For Two Years From The Date Of Final Acceptance
Automatic Voltage Stabilizer Technical Parameter	Outer cabinet should been annealed after proper anti rust process	Yes
	LED Indication	Mains ON

विवरण/Specification	विशिष्ट का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>Automatic Voltage Stabilizer Features</b>	Over load protection at input, automatic cut-off time between 2-4 seconds of Under voltage/Over voltage cut off at output, Maximum power loss due to leakage of current should be less than 2%, Restart delay of 6-9 minutes after the cut-off along with quick start button, The output voltage should be stable and capable of operating continuously with respect to any change in ambient temperature and regulation should be automatic, Transformers copper wound on CRGO laminations, as per ISO specification for high efficiency and low loss, Voltmeter (digital/analog) with selector switch to indicate input and output voltage with off-position, The output side shall be provided With C series MCB with breaking current of 10KA duly ISI marked, Provision for two output sockets of 15Amps ISI marked, Details of transformer should be furnished, such as rating, weight of winding and core, which should be verified during inspection, A Metal Oxide Varistor (MOV) for surge suppression should be provided and connected directly on the power line
	Protection	Over Voltage, Under Voltage, Over Load, Short Circuit
	Input Voltage Range	220.0, 240.0 <b>(Volt)</b>
	Operational Requirement	Should be able to provide stable AC power output for a wide range of fluctuations as per required output, The output capacity of the stabilizer should be 1KVA.
	Relative Humidity	15-90 Percent
	Warranty	The warranty shall remain valid for two years from the date of final acceptance
Test Report	Availability of Test report From Central Govt/NABL/ILAC accredited lab to prove conformity to specification	Yes
	Test Report to be Furnished to the buyer on Demand	Yes

**Additional Specification Parameters - Deep Freezer For Universal Immunization Program ( 15 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Technical Specifications: 1	1. Gross Volume: 105-125 liters. 2 Construction: 2.1 Internal: Stainless 304 grade steel 2.2 External: Corrosion Resistance 3. Should have foam pad cover on top of the basket (applicable for chest type). 4. Solid door with lock and handle. 5. Type -Chest Type

Specification Parameter Name	Bid Requirement (Allowed Values)
Technical Specifications: 2	1 Type: Compression Cycled, CFC-Free (both for refrigeration and insulation), all system tubing (suction tube, freezer tube and condensing tube) should be of minimum 99.97% of pure copper coil. 2. Temperature of compartment to be maintained between (-) 25°C to (-) 15°C continuous availability of energy as recommended by WHO PQS norms for hot zone. 3. Inlet of Capillary should be outside the PUF body.
Technical Specifications: 3	1. ON/OFF Switch and power indicator should be available. 2. Refrigerant: CFC-Free 3. A Micro-processor-based control unit should be provided for setting of temperature and display following features: 4. digit digital display (to one decimal point) of cabinet temperature. The sensor should be placed 25 to 50 mm above base of chamber.
Technical Specifications: 4	1 Power on LED/LCD indicator 2 Audio (minimum 65dBA) and visual alarm against the violation of set temperature range. 3 Min & Max cabinet temperature digital display of last 24 hrs and breaches during last 24 hrs 4 Temperature manual control with one decimal degree scale. 5 The unit should be sealed protected from dust, moisture or condensed water falling over it. 6 Sensor accuracy for digital controller $\pm 0.5^{\circ}\text{C}$ .
Standards and Safety:	1. Should meet WHO/UNICEF Standard WHO/POS/E03/FZ03.1.for Deep Freezers. 2. Test and inspection as per WHO procedure reference WHO/PQS/E03/FZ03-VP.1 testing should be carried out from WHO certified lab/NABL/STOC Labs. 3 Colour code: Blue

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Shantaben Rameshbhai Garasiya	382016,Health department, District Panchayat, Sector-17, Gandhinagar, Gujarat	15	15

#### क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

##### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

##### 2. Generic

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.



### 3. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

#### Technical Specifications

##### Deep Freezer (Small)

### **1. Description of Function:**

- 1.1 Deep freezers maintain temperatures (-) 25°C to (-) 15°C, to store vaccines and freeze ice packs,
- 1.2 Deep Freezers are required at various level of health facilities to store vaccine and prepare ice packs required for passive cooling in vaccines carriers and cold boxes.
- 1.3 Equipment only work as a Deep Freezer to maintain temperatures (-)25°C to (-) 15°C, not as a combo.

### **2. Operational Requirements:**

- 2.1 Target holdover time should be not less than 2hrs and 30 minutes.
- 2.2 Hot and cold compressor starting at 172 volts (22% below rated voltage).
- 2.3 Manufacturing process of the product should not use or produce hazardous chemicals-gases.
- 2.4 Provision for drainage for the waste water.
- 2.5 Should have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor.
- 2.6 The unit should have ground clearance minimum 100 mm.

### **3. Technical Specifications:**

- 3.1 Gross Volume: 105-125 liters.
- 3.2 Construction:
  - 3.2.1 Internal: Stainless 304 grade steel
  - 3.2.2 External: Corrosion Resistance
- 3.3 Should have foam pad cover on top of the basket (applicable for chest type).
- 3.4 Solid door with lock and handle.
- 3.5 **Type:** Compression Cycled, CFC-Free (both for refrigeration and insulation), all system tubing (suction tube, freezer tube and condensing tube) should be of minimum 99.97% of pure copper coil.
- 3.6 Temperature of compartment to be maintained between (-) 25°C to (-) 15°C continuous availability of energy as recommended by WHO PQS norms for hot zone.
- 3.7 Inlet of Capillary should be outside the PUF body.
- 3.8 ON/OFF Switch and power indicator should be available.
- 3.9 Refrigerant: CFC-Free

3.10 A Micro-processor-based control unit should be provided for setting of temperature and display following features:

3.10.1 3-digit digital display (to one decimal point) of cabinet temperature. The sensor should be placed 25 to 50 mm above base of chamber.

3.10.2 Power on LED/LCD indicator

3.10.3 Audio (minimum 65dBA) and visual alarm against the violation of set temperature range.

3.10.4 Min & Max cabinet temperature digital display of last 24 hrs and breaches during last 24 hrs

3.10.5 Temperature manual control with one decimal degree scale.

3.10.6 The unit should be sealed protected from dust, moisture or condensed water falling over it.

3.10.7 Sensor accuracy for digital controller  $\pm 0.5^{\circ}\text{C}$ .

## **4. Micro-processor System Configuration:**

4.1 Programmable Micro-processor control unit with child lock facility.

4.2 Should have provision to set minimum and maximum temperature at  $0.1^{\circ}\text{C}$  to programme the unit for continuous operation.

4.3 Should have provision for defrosting program.

## **5. Accessories, spares and warranty:**

5.1 The equipment should have minimum warranty including comprehensive maintenance of sixty months after installation or sixty-six months after the supply whichever is later.

5.2 Vaccine Storage Basket allowing free circulation of air, having the size to be able to accommodate 4 to 6 of them in the unit and suitable to match the net volume requirement.

5.3 Stem alcohol thermometer (specifications and standard as per MOHFW approved Annexure-1)-Two pieces per unit.

5.4 The supplier is required to maintain all the spare parts throughout the warranty period and not less than ten years.

## **6. Environmental factors:**

6.1 The unit shall be capable of operating continuously in ambient temperature as per WHO PQS norms for hot zone ( $+43^{\circ}\text{C}$ ).

## **7. Power Supply:**

7.1 Power input to be 220-240V AC, 50Hz suitable as per Indian standard (The plug should be flexible and unbreakable sealed rubber type).

7.2 Voltage stabilizer as per the MoHFW approved specifications and standard enclosed as Annexure-2 (applicable for external stabilizer)

## **8. Standards and Safety:**

8.1 Should meet WHO/UNICEF Standard WHO/POS/E03/FZ03.1 for Deep Freezers.

8.2 Test and inspection as per WHO procedure reference WHO/PQS/E03/FZ03-VP.1 testing should be carried out from WHO certified lab/NABL/STOC Labs.

8.3 Colour code: Blue

## **9. Documentation:**

9.1 A paper copy of user/operator manuals to be supplied in English.

9.2 A paper copy of technical/wiring diagram/maintenance manuals to be supplied in English.

9.3 Certificate of inspection for technical compliance from an independent laboratory approved/recognized by WHO certified/National Accreditation Board for Laboratories/STQC Labs is essential. Certificate of testing should be currently valid till supply and same must be verified by inspecting authority.

9.4 List of important spare parts and accessories with their part number and costing.

## **10. Packing of the equipment during shipment:**

10.1 The supplier should provide strong and sufficient packing to ensure safe arrival of goods at the destination free from loss or damage.

10.2 A vertical arrow should be marked at the all sides of packages to ensure transportation of equipment in vertical position. TOP and BOTTOM should also be written.

10.3 To put label and signage's for HANDLE WITH CARE ON ALL SIDES OF THE CRATES as per packing & shipment norms.

## **11. Following messages should be written in appropriate visibility at the Top/Front of the DEEP FREEZER:**

11.1 Place Deep Freezer at least 10cms away from the wall and 20cms away from other equipment for free air circulation.

11.2 Use voltage stabilizer provided for the DEEP FREEZER.

11.3 Safe temperature range (-)25°C to (-)15°C.

11.4 At CHC/PHC use deep freezer for freezing of ice packs only.

11.5 Open the lid, only when needed.

11.6 Store only Polio vaccines at Regional/District store.

11.7 Keep all vaccine in wire baskets provided.

- 11.8 Leave space between the vaccine boxes and ice packs for air circulation.
- 11.9 Place a thermometer at the place provided for vaccine.
- 11.10 Avoid removing thermometer from the unit while reading temperature.
- 11.11 Net vaccine storage capacity in Litres.
- 11.12 Hold over time in hrs.

#### 4. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

### अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer

is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**