

**बिड दस्तावेज़ / Bid Document**

| बिड विवरण / Bid Details   |  |
|---|--|
| बिड बंद होने की तारीख/समय / Bid End Date/Time   | 13-12-2025 17:00:00  |
| बिड खुलने की तारीख/समय / Bid Opening Date/Time  | 13-12-2025 17:30:00  |
| बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)   | 180 (Days)   |
| मंत्रालय/राज्य का नाम / Ministry/State Name   | Gujarat  |
| विभाग का नाम / Department Name  | Panchayats And Rural Housing Department Gujarat  |
| संगठन का नाम / Organisation Name  | Gandhinagar District Panchayat   |
| कार्यालय का नाम / Office Name   | Health   |
| कुल मात्रा / Total Quantity   | 5  |
| वस्तु श्रेणी / Item Category  | Ice Lined Refrigerator (Q3)  |
| बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)             | 1 Lakh (s)   |
| मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) / OEM Average Turnover (Last 3 Years)                                    | 1 Lakh (s)   |
| उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service            | 3 Year (s)   |
| एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / MSE Exemption for Years of Experience and Turnover       | No   |
| स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Exemption for Years of Experience and Turnover | No   |
| विक्रेता से मांगे गए दस्तावेज़ / Document required from seller  | Experience Criteria, Past Performance, Bidder Turnover, OEM Authorization Certificate, OEM Annual Turnover<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |

| बिड विवरण/Bid Details  |  |
|--|--|
| क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid? | Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)   |
| बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension  | 1  |
| दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended  | 3  |
| ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count  | 1  |
| विगत प्रदर्शन /Past Performance  | 10 %   |
| बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled   | Yes  |
| रिवर्स नीलामी योग्यता नियम/RA Qualification Rule   | 50% Lowest Priced Technically Qualified Bidders  |
| बिड का प्रकार/Type of Bid  | Two Packet Bid   |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation  | 2 Days   |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)                                     | No   |
| अनुमानित बिड मूल्य /Estimated Bid Value  | 250500   |
| Payment Timelines  | Payments shall be made to the Seller within <b>45</b> days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC) |
| मूल्यांकन पद्धति/Evaluation Method   | Total value wise evaluation  |
| मध्यस्थता खंड/Arbitration Clause   | No   |
| सुलह खंड/Mediation Clause  | No   |

#### ईएमडी विवरण/EMD Detail

|                   |    |
|-------------------|----|
| आवश्यकता/Required | No |
|-------------------|----|

#### ईपीबीजी विवरण /ePBG Detail

|  |                     |
|--|---------------------|
| एडवाइजरी बैंक/Advisory Bank  | State Bank of India |
| ईपीबीजी प्रतिशत (%) / ePBG Percentage (%)                          | 0.01                |
| ईपीबीजी की आवश्यक अवधि (माह) / Duration of ePBG required (Months). | 6                   |

(a). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

#### लाभार्थी / Beneficiary :

Ao  
Health, Panchayats and Rural Housing Department Gujarat, Gandhinagar District Panchayat,  
(Shantaben Rameshbhai Garasiya)

#### UIN Number NCTGC2415P

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमआईआई खरीद वरीयता/MII Purchase Preference

|  |    |
|--|----|
| एमआईआई खरीद वरीयता/MII Purchase Preference | No |
|--|----|

#### एमएसई खरीद वरीयता/MSE Purchase Preference

|  |     |
|--|-----|
| एमएसई खरीद वरीयता/MSE Purchase Preference  | Yes |
| सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within $L1+X\%$ | 15  |
| सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference                                | 25  |

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a

certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25 % percentage of total quantity. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023. [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

6. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

7. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

## Ice Lined Refrigerator ( 5 pieces )

### तकनीकी विशिष्टियाँ /Technical Specifications

[\\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

| विवरण/Specification | विशिष्टि का नाम /Specification Name    | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|--|---|
| Construction        | Type of Door (with lock and handle)    | Chest Door  |
|                     | Material of Construction of body       | Prepainted Galvanised Steel                                     |
|                     | Net Vaccine Storage Capacity (minimum) | 200.0 - 500.0 Or higher (liter)                                 |

| विवरण/Specification | विशिष्टि का नाम /Specification Name   | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)  |
|---------------------|---|--|
| Other               | Additional Features   | Corrosion resistant material, Appropriate Provision for drainage for the waste water, Optimized refrigeration system design, CFC-free high density foam insulation, Specially Integrated Icelineing, Green based technology, having recyclable components., Temperature stability provisions, having ballast boxes on the top, Equipped with QR code to work on unit tracker APP of the manufacturer |
| Warranty            | Warranty  | 5.0 (year)   |
| Test Reports        | Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification | Yes  |
|                     | Test Report to be furnished to the buyer on demand  | Yes  |

#### Additional Specification Parameters - Ice Lined Refrigerator ( 5 pieces )

| Specification Parameter Name | Bid Requirement (Allowed Values)  |
|------------------------------|---|
| Technical Specifications: 1  | 1. Net Vaccine Storage Capacity: 180-250 liters within basket in place. 2. Freeze Protection: Grade-A, User Independent Freeze Protection 3 Corrosion Resistance Body External as well as Internal 4 Chest door type with CFC-free insulation 5 Should have horizontal water cool pack covering the top of the basket in Chest type equipment. In case of front opening equipment appropriate provision to reduce cooling loss at the time of door opening. |
| Technical Specifications: 2  | 3 Temperature of a full vaccines to remain 2 deg * C to + 8 deg * C during continuous availability of energy at ambient temperature as per WHO PQS Standard (WHO/PQS/E03/RF03.1.for Ice Lined Refrigerators) with intermittent/ continuous electricity supply 8 hrs. in a 24 hrs. cycle. The temperature difference between any two points in the cabinet should not be more than 2 deg * C once stabilized.  |
| Technical Specifications: 3  | 1. Inlet of Capillary should be outside the PUF body. 2. A Microprocessor based control unit should be provided for setting of temperature and display following features: 3. 3-digit digital display (to one decimal point) of cabinet temperature. 4 Type - Chest Type  |
| Technical Specifications: 4  | 1 Solid door with lock and handle. 1.1 Power on LED/LCD indicator 2. Audio (minimum 65 dBA) and visual alarm against the violation of temperature range less than 2 deg * C and more than + 8 deg * C ) 3. The unit should be sealed/protected from dust, moisture or condensed water falling over it 4.Sensor accuracy for digital controller $\pm 0.5^{\circ}\text{C}$  |
| Standards and Safety: 5      | 1. Should meet WHO-PQS Standard (WHO/PQS/E03/RF03.1. for Ice Lined Refrigerators) 2. Test and inspection as per WHO-PQS procedure reference WHO/PQS/E03/RF03-VP.1 Testing should be carried out from WHO certified lab/NABL/STQC Labs. Certificate of testing should be currently valid till the supply and same must be verified by inspecting authority. 3. Colour code: WHITE  |

\* Bidders offering must also comply with the additional specification parameters mentioned above.

## अतिरिक्त विशिष्टि दस्तावेज़ /Additional Specification Documents

|                                   |                      |
|-----------------------------------|----------------------|
| Applicable Specification Document | <a href="#">View</a> |
|-----------------------------------|----------------------|

## परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

| क्र.सं./S.N<br>o. | परेषिती/रिपोर्टिंग<br>अधिकारी /Consignee<br>Reporting/Officer | पता/Address   | मात्रा /Quantity | डिलीवरी के दिन/Delivery<br>Days |
|-------------------|---|---|------------------|---------------------------------|
| 1                 | Shantaben<br>Rameshbhai<br>Garasiya                           | 382016,Health department,<br>District Panchayat, Sector-17,<br>Gandhinagar, Gujarat | 5                | 15                              |

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

### 2. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

## Technical Specifications Ice Line Refrigerator (Large)

### 1. Description of Function:

1.1 Ice-lined refrigerators maintain temperatures of  $2^{\circ}\text{C}$  to  $8^{\circ}\text{C}$  Minimum 8 hrs Continuous or intermittent power supply should be sufficient per 24 hrs to maintain vaccine temperature between  $2^{\circ}\text{C}$  to  $-8^{\circ}\text{C}$

1.2 Ice-lined refrigerators are required at various level of health facilities since electricity supplies are rarely perfect and standby electricity supplies may not be available.

### 2. Operational Requirements:

2.1 Vaccine storage is required for Routine Immunization, Campaign and new vaccine introduction.

- 2.2 Target holdover time should be 24 hrs or more in a continuous external temperature of 43 deg \* C
- 2.3 Hot and cold compressor starting at 172 volts (22% below rated voltage).
- 2.4 Manufacturing process of the product should not use or produce hazardous Chemicals-gases.
- 2.5 Appropriate Provision for drainage for the waste water.
- 2.6 Should have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor.
- 2.7 The unit should have ground clearance of minimum 100mm.

### **3. Technical Specifications:**

- 3.1 Net Vaccine Storage Capacity: 180-250 liters within basket in place.
- 3.2 Freeze Protection: Grade-A, User Independent Freeze Protection
- 3.3 Construction: Internal: Stainless 304 grade steel/Corrosion Resistant Polymer
- 3.4 Corrosion Resistance Body External as well as Internal
- 3.5 Chest door type with CFC- free insulation
- 3.6 Should have horizontal water cool pack covering the top of the basket in Chest type equipment. In case of front opening equipment appropriate provision to reduce cooling loss at the time of door opening.
- 3.7 Solid door with lock and handle.
- 3.8 **Type:** Compression Cycled, CFC-Fuoco (both for refrigeration and insulation) All system tubing (suction tube, freezer tube and condensing tube) should be of minimum 99.97% of pure copper coil.
- 3.9 Refrigerant: CFC-Free
- 3.10 Temperature of a full vaccines to remain 2 deg \* C to + 8 deg \* C during continuous availability of energy at ambient temperature as per WHO PQS Standard (WHO/PQS/E03/RF03.1.for Ice Lined Refrigerators) with intermittent/ continuous electricity supply 8 hrs. in a 24 hrs. cycle. The temperature difference between any two points in the cabinet should not be more than 2 deg \* C once stabilized.
- 3.11 Inlet of Capillary should be outside the PUF body.
- 3.12 ON/OFF Switch and power indicator should be available
- 3.13 A Microprocessor based control unit should be provided for setting of temperature and display following features:
  - 3.13.1 3-digit digital display (to one decimal point) of cabinet temperature.
  - 3.13.2 Power on LED/LCD indicator
  - 3.13.3 Audio (minimum 65 dBA) and visual alarm against the violation of temperature range less than 2 deg \* C and more than + 8 deg \* C )
  - 3.13.4 The unit should be sealed/protected from dust, moisture or condensed water falling over it
  - 3.13.5 Sensor accuracy for digital controller  $\pm 0.5^{\circ}\text{C}$

### **4. Micro-processor System Configuration:**

- 4.1 Programmable Micro-processor control unit with child lock facility.
- 4.2 Should have provision to set minimum and maximum temperature at 0.1 °C to programme the unit for continuous operation.

## **5. Accessories, spares and warrantee:**

- 5.1 The equipment should have minimum warrantee including comprehensive maintenance of sixty months after installation or sixty-six months after the supply whichever is later.
- 5.2 Vaccine Storage Basket/Tray allowing free circulation of air, having the size to be able to accommodate 4 to 6 of them in the unit and suitable to match the net volume requirement
- 5.3 Stem Alcohol thermometer (specifications and standard as per MoHFW approved Annexure-1)-two pieces per unit.
- 5.4 The supplier is required to maintain all the spare parts throughout the warrantee period and not less than ten years.

## **6. Environmental factors:**

- 6.1 The unit shall be capable of operating continuously in ambient temperature as per WHO-PQS norms for hot zone (+43°C).

## **7. Power Supply:**

- 7.1 Power input to be 220-240V AC, 50Hz suitable as per Indian standard (The plug should be flexible and unbreakable sealed rubber type)
- 7.2 Voltage stabilizer as per the MOHFW approved specifications and standard enclosed as Annexure-2 (applicable for external stabilizer)

## **8. Standards and Safety:**

—

- 8.1 Should meet WHO-PQS Standard (WHO/PQS/E03/RF03.1. for Ice Lined Refrigerators)
- 8.2 Test and inspection as per WHO-PQS procedure reference WHO/PQS/E03/RF03-VP.1 Testing should be carried out from WHO certified lab/NABL/STQC Labs. Certificate of testing should be currently valid till the supply and same must be verified by inspecting authority.
- 8.4 Colour code: WHITE

## **9. Documentation:**

- 9.1 A paper copy of user/operator manuals to be supplied in English.
- 9.2 A paper copy of technical/wiring diagram/maintenance manuals to be supplied in English.
- 9.3 Certificate of inspection for technical compliance from an independent laboratory approved/recognized by WHO certified/National Accreditation Board for Laboratories/STQC Lab is essential.
- 9.4 List of important spare parts and accessories with their part number and costing.



## **10 .Packing of the equipment during shipment:**

10.1 The supplier should provide strong and sufficient packing to ensure safe arrival of goods at the destination free from loss or damage.

10.2 A vertical arrow should be marked at the all sides of packages to ensure transportation of equipment in vertical position. TOP and BOTTOM should also be written.

10.3 To put label and signage's for HANDLE WITH CARE ON ALL SIDES OF THE CRATES as per packing & shipment norms.

## **11. Following messages/information should be written in appropriate visibility at the Top/Front of the ILR:**

11.1 Place Ice-lined refrigerator at least 10cms away from the wall and 20cms away from other equipment for free air circulation.

11.2 Use voltage stabilizer provided with the ILR. In case of an inbuilt stabilizer, it should be mentioned on the equipment "This equipment has an inbuilt voltage stabilizer".

11.3 Safe temperature range +2°C to +8°C

11.4 Store all UIP vaccines in ILR at CHC/PHC (OPV should be stored in deep freezer at State/Regional and district vaccine store).

11.5 Open the lid/door, only when needed.

11.6 Store only UIP vaccines.

11.7 Leave space between the vaccine boxes for air circulation.

11.8 Place a thermometer between the vaccines Boxes,

11.12 Net vaccine storage capacity in Litres.

11.13 Hold over time in hrs.

11.14 Serial number of the equipment.

11.15 Model number of the equipment.

11.16 Toll free number for maintenance support within warranty period.

## **Technical Specifications:**

1. Net Vaccine Storage Capacity: 180-250 liters within basket in place.
2. Freeze Protection: Grade-A, User Independent Freeze Protection
3. Construction: Internal: Stainless 304 grade steel/Corrosion Resistant Polymer
4. Corrosion Resistance Body External as well as Internal
5. Should have horizontal water cool pack covering the top of the basket in Chest type equipment. In case of front opening equipment appropriate provision to reduce cooling loss at the time of door opening.
6. Solid door with lock and handle.

**7 Type:** Compression Cycled, CFC-FuCo (both for refrigeration and insulation) All system tubing (suction tube, freezer tube and condensing tube) should be of minimum 99.97% of pure copper coil.

8 Temperature of a full vaccines to remain 2 deg \* C to + 8 deg \* C during continuous availability of energy at ambient temperature as per WHO PQS Standard (WHO/PQS/E03/RF03.1 for Ice Lined Refrigerators) with intermittent/ continuous electricity supply 8 hrs. in a 24 hrs. cycle. The temperature difference between any two points in the cabinet should not be more than 2 deg \* C once stabilized.

9 Inlet of Capillary should be outside the PUF body.

10 ON/OFF Switch and power indicator should be available

11. A Microprocessor based control unit should be provided for setting of temperature and display following features:

12 3-digit digital display (to one decimal point) of cabinet temperature.

13 Power on LED/LCD indicator

14 Audio (minimum 65 dBA) and visual alarm against the violation of temperature range less than 2 deg \* C and more than + 8 deg \* C )

15 The unit should be sealed/protected from dust, moisture or condensed water falling over it

16 Sensor accuracy for digital controller  $\pm 0.5^{\circ}\text{C}$

## **Standards and Safety:**

—

1 Should meet WHO-PQS Standard (WHO/PQS/E03/RF03.1. for Ice Lined Refrigerators)

2 Test and inspection as per WHO-PQS procedure reference WHO/PQS/E03/RF03-VP.1 Testing should be carried out from WHO certified lab/NABL/STQC Labs. Certificate of testing should be currently valid till the supply and same must be verified by inspecting authority.

3 Colour code: WHITE

### 3. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

### 4. Generic

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

PHC  
CHC  
UPHC  
OF DP Gandhinagar  
Gujarat  
.

## अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in

the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा। In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**