

**बिड दस्तावेज़ / Bid Document**

| बिड विवरण/Bid Details   |   |
|---|---|
| बिड बंद होने की तारीख/समय /Bid End Date/Time  | 13-11-2025 11:00:00   |
| बिड खुलने की तारीख/समय /Bid Opening Date/Time   | 13-11-2025 11:30:00   |
| बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)   | 180 (Days)  |
| मंत्रालय/राज्य का नाम/Ministry/State Name   | Ministry Of Defence   |
| विभाग का नाम/Department Name  | Department Of Military Affairs  |
| संगठन का नाम/Organisation Name  | Indian Army   |
| कार्यालय का नाम/Office Name   | *****   |
| कुल मात्रा/Total Quantity   | 1   |
| वस्तु श्रेणी /Item Category   | Echo Cardiography Scanning Machine (V2) (Q2)  |
| उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service  | 3 Year (s)  |
| वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Exemption for Years Of Experience and Turnover   | Yes   Complete  |
| स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Exemption for Years Of Experience and Turnover   | Yes   Complete  |
| विक्रेता से मांगे गए दस्तावेज़/Document required from seller  | Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Compliance of BoQ specification and supporting document<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid? | Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)  |

| बिड विवरण/Bid Details  |                             |
|--|-----------------------------|
| बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / <b>Minimum number of bids required to disable automatic bid extension</b>   | 2                           |
| दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / <b>Number of days for which Bid would be auto-extended</b>   | 3                           |
| ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / <b>Number of Auto Extension count</b>   | 1                           |
| विगत प्रदर्शन / <b>Past Performance</b>  | 80 %                        |
| बिड से रिवर्स नीलामी सक्रिय किया/ <b>Bid to RA enabled</b>   | No                          |
| बिड का प्रकार/ <b>Type of Bid</b>  | Two Packet Bid              |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय / <b>Time allowed for Technical Clarifications during technical evaluation</b>  | 2 Days                      |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/ <b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b> | No                          |
| मूल्यांकन पद्धति/ <b>Evaluation Method</b>   | Total value wise evaluation |
| मध्यस्थता खंड/ <b>Arbitration Clause</b>   | No                          |
| सुलह खंड/ <b>Mediation Clause</b>  | No                          |

**ईएमडी विवरण/EMD Detail**

|                   |    |
|-------------------|----|
| आवश्यकता/Required | No |
|-------------------|----|

**ईपीबीजी विवरण /ePBG Detail**

|                   |    |
|-------------------|----|
| आवश्यकता/Required | No |
|-------------------|----|

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

**एमआईआई के लिए आरक्षित / Reserved for Make In India products**

|   |     |
|---|-----|
| एमआईआई के लिए आरक्षित / Reserved for Make In India products | Yes |
|---|-----|

**एमएसई के लिए आरक्षित / Reserved for MSE**

|  |  |
|--|--|
|  |  |
|--|--|

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking exemption from Experience Criteria, shall upload the supporting documents to prove his eligibility for exemption.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking exemption from Turnover, shall upload the supporting documents to prove his eligibility for exemption.
3. If the bidder is a DPIIT registered Startup, the bidder shall be exempted from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking exemption from Experience Criteria, shall upload the supporting documents to prove his eligibility for exemption.
4. If the bidder is a DPIIT registered Startup, the bidder shall be exempted from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking exemption from Turnover shall upload the supporting documents to prove his eligibility for exemption.
5. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
6. Bid reserved for Make In India products: : Procurement under this bid is reserved for purchase from Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. However, eligible micro and small enterprises will be allowed to participate. The minimum local content to qualify as a class 1 local supplier is denoted in the bid document. All bidders must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which the bid is liable to be rejected. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020 . In case Buyer has selected Purchase preference to Micro and Small Enterprises clause in the bid, the same will get precedence over this clause.
7. Procurement under this bid is reserved for purchase from Micro and Small Enterprises having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal. If the bidder wants to avail themselves of the reservation benefit, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible to participate in this bid. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility based on documentary evidence submitted, while evaluating the bid. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.
9. Short Duration Bid has been published by the Buyer with the approval of the Competent authority due to Emergency procurement of critical products/services.

### **Echo Cardiography Scanning Machine (V2) ( 1 pieces )**

**(Minimum 50% Local Content required for qualifying as Class 1 Local Supplier)**

## तकनीकी विशिष्टियाँ /Technical Specifications

\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

| विवरण/Specification        | विशिष्टि का नाम /Specification Name                              | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)                      |
|----------------------------|--|--|
| Types of Probes & Features | <b>Cardiac and Vascular Probes Configuration</b>                 | TTE Phased Array, TTE Matrix Array, TEE Phased Array, TEE Matrix Array, Linear Probe |
|                            | <b>TTE(Transthoracic) Probe Type</b>                             | Adult, Pediatric, Neonatal   |
|                            | <b>TEE Adult Probe biplane and Multiplane frequency, in MHz</b>  | 2 to 7.5 MHz ( $\pm 2$ )   |
|                            | <b>TEE Pediatric Probe Required</b>                              | No   |
|                            | <b>TEE Pediatric Probe frequency, in MHz</b>                     | Not Applicable   |
|                            | <b>Linear Probe for linear/vascular applications required</b>    | Yes  |
|                            | <b>Adult/Pediatric Linear probe frequency in MHz</b>             | 5 to 14 MHz ( $\pm 2$ )  |
|                            | <b>No of Adult TTE Probes Required</b>                           | 1  |
|                            | <b>No of Pediatric TTE Probes Required</b>                       | NA   |
|                            | <b>No of Neonatal TTE Probes Required</b>                        | NA   |
|                            | <b>No of Adult TEE Probes Required</b>                           | NA   |
|                            | <b>No of Pediatric TEE Probes Required</b>                       | NA   |
|                            | <b>No of Pediatric 3D/4D TEE Probes Required</b>                 | NA   |
|                            | <b>No of Neonatal TEE Probes Required</b>                        | NA   |
|                            | <b>No of Adult/pediatric Linear Probes supplied</b>              | NA   |
|                            | <b>Dedicated 3D/4D volume imaging for cardiology TTE and TEE</b> | No   |
|                            | <b>Live 3 D imaging in color doppler system</b>                  | No   |

| विवरण/Specification                   | विशिष्टि का नाम /Specification Name   | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)                    |
|---------------------------------------|---|--|
|                                       | <b>Colour Doppler system with all application packages ,Quad loop for serial studies with high frame rate revies, harmonic imaging capability in all modes and integrated stress echo package digital storage and retrieval</b> | No   |
|                                       | <b>Provision of Tissue Colorization (B-Colour) for improved contrast resolution</b>   | No   |
|                                       | <b>Application software</b>   | Adult, Pediatric, Neonatal, Vascular applications and Transesophageal applications |
|                                       | <b>Cine Loop Features</b>   | Yes  |
|                                       | <b>Frame grabber facility for post analysis</b>   | No   |
|                                       | <b>Facility of Various maps for pre and post processing</b>   | No   |
|                                       | <b>ECG trigger facility</b>   | No   |
|                                       | <b>User defined Preset</b>  | No   |
|                                       | <b>Dedicated integrated dynamic stress echo package for flexible user defined protocols with stacked sub loops facility and contrast stress protocol usable for stress echocardiography</b>                                     | No   |
|                                       | <b>Number of transducer/probe ports</b>   | 1  |
| Scan Modes and Performance Parameters | <b>Cardiac 3-D freehand Option</b>  | No   |
|                                       | <b>Cardiac 3D/4D automatic Option</b>   | No   |
|                                       | <b>Contrast Harmonic/Agent Imaging with angio mode</b>  | No   |
|                                       | <b>Color Doppler 3-D/4-D option</b>   | No   |
|                                       | <b>Power Dopple 3-D/4-D option</b>  | No   |
|                                       | <b>Contrast-enhanced ultrasound (CEUS) option</b>   | Yes  |
|                                       | <b>Tissue synchronization, tissue velocity</b>  | Yes  |

| विवरण/Specification          | विशिष्टि का नाम /Specification Name                                      | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)  |
|------------------------------|--|--|
|                              | Cardiac Shear wave elastography  | No   |
|                              | Cardiac compression elastography   | No   |
|                              | Provision for higher sensitivity for low frequency doppler in all probes | Yes  |
|                              | Latest generation Electronic Phased/sector array Colour Doppler system   | Yes  |
|                              | Number of Channels in color doppler system                               | 500000 or more (if channels are being designated as digital channels)  |
|                              | System should be capable of generating real time live 3- D images        | No   |
| Image display and processing | DICOM 3 Compliant  | Yes  |
|                              | Automated B-mode (2-D) image   | Yes  |
|                              | Automated CDI image optimization   | Yes  |
|                              | Automated PW Doppler image optimization                                  | Yes  |
|                              | Touch Screen   | Yes  |
|                              | Monitor display size, cm (in)  | 21 Inch( $\pm 1$ ) Or higher   |
|                              | Split screen   | Twin View  |
|                              | Single/dual monitors   | Single   |
|                              | Speckle-tracking strain and strain rate                                  | No   |
|                              | Analysis and calculation packages  | 2-D wall motion tracking, 3-D wall motion tracking, Cardiac, Mitral valve analysis, User-programmable calculations, No |
|                              | Type of printer provided with the echo cardio machine                    | Color Laser Printer  |
|                              | Type of processor in PC  | NA   |
|                              | Conectivity between Echo cardiogram and PC                               | NA   |
|                              | PC with all software inclusive interfaced with echocardiograph machine   | NA   |
|                              | Inkjet printer provided with computer for reporting                      | No   |

| विवरण/Specification | विशिष्टि का नाम /Specification Name       | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|---|---|
|                     | Glossy Colour Print Paper supplied        | Not provided  |
|                     | ECG Cable                                 | NA  |
|                     | Echo cardiography sytem to be DICOM Ready | Yes   |
|                     | Dicom Interfacability to HIS              | Yes   |
|                     | Dicom Interfacability Dicom to PACS       | Yes   |
|                     | Interfacability Dicom to RIS              | Yes   |

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

| क्र.सं./S.No. | परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer | पता/Address | मात्रा /Quantity | डिलीवरी के दिन/Delivery Days |
|---------------|---|-------------|------------------|------------------------------|
| 1             | *****   | *****Shimla | 1                | 15                           |

**Special terms and conditions-Version:1 effective from 15-04-2024 for category Echo Cardiography Scanning Machine (V2)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty

shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.



## 2. OEM

**IMPORTED PRODUCTS:** In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

## 3. Inspection

**Nominated Inspection Agency:** On behalf of the Buyer organization, any one of the following Inspection Agency would be conducting inspection of stores before acceptance:  
Pre-dispatch Inspection at Seller Premises (applicable only if pre-dispatch inspection clause has been selected in ATC):

Seller

Post Receipt Inspection at consignee site before acceptance of stores:  
BOO

## 4. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

## 5. Warranty

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

## 6. Turnover

**Bidder Turn Over Criteria:** The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

## 7. Turnover

**OEM Turn Over Criteria:** The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

## 8. Past Project Experience

**Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed.b. Execution certificate by client with contract value.c. Any other document in support of contract execution like Third Party Inspection release note, etc.**Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be

considered as valid proof for meeting the experience criteria: a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed. b. Execution certificate by client with contract value. c. Any other document in support of contract execution like Third Party Inspection release note, etc.

## 9. Past Project Experience

The Bidder / OEM {themselves or through reseller(s)}, should have executed project for supply and installation / commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:

- (i) Single order of at least 35% of estimated bid value; or
- (ii) Two orders of at least 20% each of estimated bid value; or
- (iii) Three orders of at least 15% each of estimated bid value.

Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion

## अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**