

**बिड दस्तावेज़ / Bid Document**

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	04-12-2025 14:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	04-12-2025 14:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	Assam
विभाग का नाम / Department Name	Health And Family Welfare Department Assam
संगठन का नाम / Organisation Name	Assam Medical Services Corporation Limited
कार्यालय का नाम / Office Name	Narengi
कुल मात्रा / Total Quantity	82
वस्तु श्रेणी / Item Category	Customised Ambulance with Equipment (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	1463 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) / OEM Average Turnover (Last 3 Years)	11709 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service	3 Year (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / MSE Exemption for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Exemption for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़ / Document required from seller	Experience Criteria, Past Performance, Bidder Turnover, Certificate (Requested in ATC), OEM Authorization Certificate, OEM Annual Turnover, Additional Doc 1 (Requested in ATC), Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

बिड विवरण/Bid Details	
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	80 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	Customised Ambulance with Equipment
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य /Estimated Bid Value	292726400
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

#### ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईएमडी राशि/EMD Amount	5854528

#### ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
-----------------------------	-----------

ईपीबीजी प्रतिशत (%) / ePBG Percentage (%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) / Duration of ePBG required (Months).	38

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

#### लाभार्थी / Beneficiary :

Managing Director

Office of the Managing Director, Assam Medical Services Corporation Limited, Central Drug Ware House Campus, Patherquery, Narengi, Guwahati, Assam-781026

(Dr. Lakshmanan S, Ias)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

#### एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within $L1+X\%$	20
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	25

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value

should meet this criterion.

3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Purchase preference to Micro and Small Enterprises (MSEs) from the State of Bid Inviting Authority : Purchase preference will be given to MSEs as Micro and Small Enterprises from the State of Bid inviting Authority. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 20% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 25 % of total quantity as defined/ decided in relevant policy.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

6. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

7. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- If number of technically qualified bidders are only 2 or 3.
- If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

#### Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
29-11-2025 15:00:00	Office of the Managing Director, Assam Medical Services Corporation Limited, Central Drug Ware House Campus, Patherquerry, Narengi, Guwahati, Assam-781026

#### Customised Ambulance With Equipment ( 42 pieces )

##### तकनीकी विशिष्टियाँ /Technical Specifications

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Generic Parameters	<b>Type of Ambulance Vehicle [as per clause 3*3 of AIS-125 (Part 1)]</b>	Type D - Road Ambulance: Advanced Life Support Ambulance

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>Category of Vehicle</b>	M2 (GVW <=5 tonnes)
	<b>Type of Wheel Drive</b>	Two wheel drive
	<b>Type of Fuel</b>	Diesel
Engine and Transmission	<b>Maximum Engine Output / Power (kW) at rated RPM</b>	75.0 - 200.0 Or higher ( <b>kiloWatt</b> )
Vehicle Parameters	<b>Maximum Gross Vehicle Weight (Kerb weight + Payload)</b>	3495.0 - 7000.0 Or higher ( <b>kilogram</b> )
Accessories/Features	<b>Air Conditioning</b>	Whole Ambulance
	GPS Based Navigation System	Yes
Ambulance Body	Number of Attendant Seats (including driver)	2
	<b>Number of Stretcher</b>	1
	<b>Features of Main Stretcher</b>	Main stretcher with an integral undercarriage
Electrical Requirement	Number of battery(ies) provided other than the vehicle battery	1, 2 Or higher
	Number of 12V connections for medical devices in patient's compartment	4, 5, 6 Or higher
	Number of 220V AC power supply connections for medical devices in patient compartment	4, 6, 8 Or higher
Warranty	<b>Vehicle Warranty Time (Years)</b>	3, 4, 5 Or higher ( <b>year</b> )
Certification Parameters	Agreed to provide medical equipment as specified in the bid with applicable OEM Warranty	Yes
	Availability of Chassis / Original Vehicle Manufacturer's authorization in case of Body Builders	Yes

**परिषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.No.	परिषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
---------------	---	-------------	------------------	------------------------------

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Sandip Dutta	781026,Central Drug Ware House, Patherquery Tiniali, Narengi, Guwahati Ph.- 9864139342	42	90

### Customised Ambulance With Equipment ( 40 pieces )

#### तकनीकी विशिष्टियाँ /Technical Specifications

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Generic Parameters	<b>Type of Ambulance Vehicle [as per clause 3*3 of AIS- 125 (Part 1)]</b>	Type C - Road Ambulance: Basic Life Support Ambulance
	<b>Category of Vehicle</b>	M2 (GVW <=5 tonnes)
	<b>Type of Wheel Drive</b>	Two wheel drive
	<b>Type of Fuel</b>	Diesel
Engine and Transmission	<b>Maximum Engine Output / Power (kW) at rated RPM</b>	75.0 - 200.0 Or higher <b>(kiloWatt)</b>
Vehicle Parameters	<b>Maximum Gross Vehicle Weight (Kerb weight + Payload)</b>	3495.0 - 7000.0 Or higher <b>(kilogram)</b>
Accessories/Features	<b>Air Conditioning</b>	Whole Ambulance
	GPS Based Navigation System	Yes
Ambulance Body	Number of Attendant Seats (including driver)	2
	<b>Number of Stretcher</b>	1
	<b>Features of Main Stretcher</b>	Main stretcher with an integral undercarriage
Electrical Requirement	Number of battery(ies) provided other than the vehicle battery	1, 2 Or higher
	Number of 12V connections for medical devices in patient's compartment	4, 5, 6 Or higher
	Number of 220V AC power supply connections for medical devices in patient compartment	4, 6, 8 Or higher
Warranty	<b>Vehicle Warranty Time (Years)</b>	3, 4, 5 Or higher <b>(year)</b>

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Certification Parameters	Agreed to provide medical equipment as specified in the bid with applicable OEM Warranty	Yes
	Availability of Chassis / Original Vehicle Manufacturer's authorization in case of Body Builders	Yes

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Sandip Dutta	781026,Central Drug Ware House, Patherquary Tiniali, Narengi, Guwahati Ph.- 9864139342	40	90

#### Special terms and conditions-Version:2 effective from 29-01-2025 for category Customised Ambulance with Equipment

##### 1. Special Terms and Conditions for Customized Ambulance with Equipment

This category is created to cater to the requirement of buyer where the buyer wants to procure Ambulance with medical devices and/or minor fabrications. This category is bid only category and Direct Purchase is not enabled.

1. The buyer shall define the equipment required in the Ambulance with complete technical specifications along with quantity. The buyer may refer AIS 125 (Part2) for list of equipment.
2. In case fabrication is required by the buyer inside the body of the Ambulance, details and drawings/pictures shall be provided by the buyer in the bid.
3. The buyer should check compliance of the vehicle to CMVR rules and compliance to AIS 125 (Part1) after fabrication and installation of medical devices.

Note for Sellers :

- For listing a catalog on GeM, the seller should specify the price for the vehicle. As the category is bid-only, the seller can submit their offer price while participating in the bid, based on the scope of work provided by the buyer.
- The seller should upload images that include the vehicle's front view, side view, and rear view with the door open.

#### क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

##### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to

increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

## 2. **Generic**

**Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

## 3. **Generic**

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 5 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

## 4. **Generic**

Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

## 5. **Generic**

**Manufacturer Authorization:** Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

## 6. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

## 7. **Turnover**

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

## 8. **Turnover**

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

## 9. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free



Telephone No. for Service Support.

#### 10. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

#### 11. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

#### 12. **Warranty**

Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

#### 13. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 7 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

#### 14. **Past Project Experience**

**Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria: a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed. b. Execution certificate by client with contract value. c. Any other document in support of contract execution like Third Party Inspection release note, etc. **Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria: a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed. b. Execution certificate by client with contract value. c. Any other document in support of contract execution like Third Party Inspection release note, etc.

#### 15. **Past Project Experience**

The Bidder / OEM {themselves or through reseller(s)}, should have executed project for supply and installation / commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:

- (i) Single order of at least 35% of estimated bid value; or
- (ii) Two orders of at least 20% each of estimated bid value; or
- (iii) Three orders of at least 15% each of estimated bid value.

Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion

#### 16. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

Assam Medical Services Corporation Limited  
payable at  
Guwahati

. Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

#### 17. **Forms of EMD and PBG**

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

Assam Medical Services Corporation Limited

. The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

#### 18. **Forms of EMD and PBG**

Bidders can also submit the EMD with Banker's Cheque in favour of

Assam Medical Services Corporation Limited  
payable at  
Guwahati

. Bidder has to upload scanned copy / proof of the BC along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

#### 19. **Forms of EMD and PBG**

Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name

Assam Medical Services Corporation Limited  
Account No.  
50100516234601  
IFSC Code  
HDFC0009584  
Bank Name  
HDFC Bank  
Branch address  
Dispur Secretariat

. Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

#### 20. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

Assam Medical Services Corporation Limited  
payable at  
Guwahati

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

#### 21. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

Assam Medical Services Corporation Limited

A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the

demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

## 22. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name

Assam Medical Services Corporation Limited

Account No.

50100516234601

IFSC Code

HDFC0009584

Bank Name

HDFC Bank

Branch address

Dispur Secretariat

. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

## 23. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

a) Specification of Medical equipment's to be supplied and installed with ALS Ambulances:

Sl	Description of the item	Specification	Qty
1	TYPE-D OXYGEN CYLINDERS WITH CYLINDER KEYS AND REGULATORS	<ol style="list-style-type: none"> <li>1. Bulk Size D- Type High pressure seamless cylinders for medical oxygen cylinder gas.</li> <li>2. Cylinders should confirm to IS: 7285, certified by the Bureau of Indian Standard (BIS) and approved by Chief Controller of Explosives (CCOE), Government of India.</li> <li>3. 46.7 Ltr. water capacity (220 Cu.Ft.), fitted with bull nose type valve as per IS:3224 and neck cap.</li> <li>4. Color code of the cylinder should be as per IS: 3933-1966 with updating till date.</li> <li>5. Certificate from Department of Explosives, Government of India should be provided for each cylinder.</li> <li>6. Working pressure should be 150 Kg. f/cm<sup>2</sup> at 15° C, Hydraulic test pressure 250 Kg.f/cm</li> <li>7. Compatible regulator should be supplied along with Oxygen Cylinder.</li> <li>8. Warranty: 1 Year or higher</li> </ol>	42
2	FLOWMETERS	<ol style="list-style-type: none"> <li>1. Back Pressure Compensated flow meter should be of accurate gas flow measurement with following feature.</li> <li>2. Control within a range of 0 – 15 LPM. (calibration within ±10%)</li> <li>3. The flow meter body should be made of brass chrome plated materials.</li> <li>4. The flow tube and shroud components should be made of clear, impact resistant polycarbonate.</li> <li>5. Warranty: 1 Year or higher</li> </ol>	42

3	HUMIDIFIERS	1. Humidifier Bottle with 1 Year Warranty.	42
4	DEFIBRILLATOR WITH RHYTHM AND PATIENT DATA RECORDING WITH CARDIAC MONITOR (ISO 60601-2-4)	<ol style="list-style-type: none"> <li>1. Biphasic, Manual and AED with voice prompt, compact and light weight</li> <li>2. Energy selection 5J to 360J in steps.</li> <li>3. Momentary energy selection access on front panel.</li> <li>4. Should have adult and paediatric paddles.</li> <li>5. Monitor should display selected and delivered energy</li> <li>6. Should have disarm facility.</li> <li>7. Energy should be delivered within 30ms after the detected R wave in synchronization mode.</li> <li>8. Charging time maximum 8 sec for 360J.</li> <li>9. Should have battery backup for 100 discharges of 360 J.</li> <li>10. Should have ECG inputs through paddles or 3 lead cables.</li> <li>11. Lead off message should appear with alert tone.</li> <li>12. Should have colour display for heart rate of size 6 inches or more.</li> <li>13. Should have alarm for high and low HR.</li> <li>14. Should have an inbuilt thermal recorder.</li> <li>15. Should have enable/disable option for printer.</li> <li>16. Should supply 2 bottles of jelly, 12 roll of thermal paper.</li> <li>17. Should supply three pairs of AED pads.</li> <li>18. Should operate on mains 230V, 50Hz.</li> <li>19. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/BIS.</li> <li>20. Should have ISO 60601-2-4 certificate.</li> <li>21. Warranty: 3 Years or higher</li> </ol>	42
5	EXTERNAL CARDIAC PACING (ISO 60601-2-4)	<ol style="list-style-type: none"> <li>1. Pacing Modes: Can be single chamber (ventricular pacing) or dual chamber (atrial and ventricular pacing) depending on the device.</li> <li>2. Pulse Duration: The length of the electrical pulse delivered, usually set between 0.1 - 1 milliseconds.</li> <li>3. Rate Range: Adjustable pacing rate to match the desired heart rhythm.</li> <li>4. Electrode Pads: Large, adhesive pads designed for optimal skin contact and minimal discomfort.</li> <li>5. ECG Monitoring: Ability to monitor the patient's ECG rhythm to verify pacing effectiveness.</li> <li>6. Power Source: Usually operates on AC power or battery depending on the model.</li> <li>7. Warranty: 3 Years or higher</li> </ol>	42
6	PORTABLE ADVANCED RESUSCITATION SYSTEM (P.A.R.S.)	<ol style="list-style-type: none"> <li>1. Portable Advanced Resuscitation System (P.A.R.S.).</li> <li>2. Warranty: 1 Year or higher</li> </ol>	42

7	NEBULIZATION APPARATUS	<ol style="list-style-type: none"> <li>1. Operating Voltage: 230 V AC</li> <li>2. Maximum pressure 35 psi.</li> <li>3. Air power: 14 liters per minute</li> <li>4. Aerosol output: 106 per minute</li> <li>5. Residual volume: 1.24 ml</li> <li>6. Droplet size: MMAD 3.3 microns</li> <li>7. Filling volume: maximum 7ml</li> <li>8. Noise level: 55 db.</li> <li>9. Provisions for fixing/Hanging in the Ambulance.</li> <li>10. Check valve to protect the device against contamination due to backward inhalation.</li> <li>11. Unbreakable lids.</li> <li>12. Dust filter.</li> <li>13. Product should be CE certified.</li> <li>14. Warranty: 1 Year or higher</li> </ol>	42
8	THORAX DRAINAGE KIT	<ol style="list-style-type: none"> <li>1. Thorax drainage kit: - 1 XRO trocar drain</li> <li>2. 1 60x50cm fenestrated drape</li> <li>3. 2 rectangular cups for detergent and disinfecting agents.</li> <li>4. 4 pads</li> <li>5. 8 10x10cm gauze dressings</li> <li>6. 2 dressing forceps</li> <li>7. 2 10ml syringes</li> <li>8. 1 38x1.2mm (18G) needle for drug aspiration</li> <li>9. 1 15x0.5mm (25G) needle for skinwheal</li> <li>10. 1 38x0.8mm (21G) needle for local anaesthesia</li> <li>11. 1 90x1.2mm (18G) needle for exploratory puncture (38mmx0.9mm (20G) for codes 639.08 and 639.14)</li> <li>12. 2 suture needles</li> <li>13. 1 scalpel</li> <li>14. 1 dual suction valve</li> <li>15. 1 plastic bag with connector for collecting the drained liquid</li> <li>16. 1 coil of adhesive bandage, 75mm wide</li> <li>17. 1 pair of gloves.</li> <li>18. Warranty: 1 Year or higher</li> </ol>	42

9	VOLUMETRIC INFUSION DEVICE	<ol style="list-style-type: none"> <li>1. Volumetric pump which can run on all commonly used IV Lines.</li> <li>2. Should have Rate, Time, Volume and Dose mode of operation.</li> <li>3. Flow rate should be of the range 1.0ml/hr to 1200 ml/hr with increment of 0.1ml/hr.</li> <li>4. Flow rate accuracy should be <math>\pm 5\%</math> or better.</li> <li>5. Presence of drop sensor is must and should have flow rate from 1 drop/min to 400 drop/min.</li> <li>6. Should have drug library of 100 or more drugs categorized in more than 15 categories.</li> <li>7. Bolus rate maximum rate should be 1200ml/hr.</li> <li>8. KVO Rate 1 ml/h for infusion flow rate <math>\leq 10</math> ml/h and 3 ml/h for infusion flow rate <math>&gt; 10</math> ml/h.</li> <li>9. Should have 3 preset level pressure mode (High, Medium, Low).</li> <li>10. Pump should display the following status Text display of current status, red alarm alert for warnings, KVO and Bolus indicator All alarms are expressed by means of light indicators, written words and sound beeps.</li> <li>11. Pump should have the following alarm options Occlusion pressure alarm, end of infusion alarm, air bubble alarm, rate error alarm and empty bag alarm, Door open alarm, low battery alarm, technical fault alarm.</li> <li>12. Should work on 220-240 V AC as well as rechargeable batteries. And battery backup should be 6 hrs or more</li> <li>13. The Display should be LCD/LED/TFT.</li> <li>14. Should have port for connectivity.</li> <li>15. Should have waterproofing of IPX3/IPX4.</li> <li>16. The device should store at least 1000 most recent infusion and alarm history events.</li> <li>17. The product should be EUCE from 4 digit notified body/USFDA/BIS certified and certificate to be submitted</li> <li>18. Manufacturer should be ISO 13485 certified.</li> <li>19. Warranty: 3 Years or higher</li> </ol>	42
10	REQUIREMENTS FOR EMERGENCY AND TRANSPORT VENTILATORS (ISO 10651-3)	<ol style="list-style-type: none"> <li>1. Compact transport ventilator pneumatic / turbine / piston technology.</li> <li>2. Suitable for Adults, Children &amp; Infants</li> <li>3. Should be able to wide range tidal volume from 50 -1500ml or better.</li> <li>4. Should have a wide frequency range 8 to 40 bpm</li> <li>5. "Models of ventilation: Volume controlled, Pressure controlled, Pressure support, Synchronized intermittent mandatory</li> <li>6. ventilation (SIMV), Assist/control mode, CPAP, Inbuilt PEEP."</li> <li>7. Alarm required: minute volume, pressure, PEEP, apnoea, occlusion, high respiration rate, disconnection</li> <li>8. Weight not more than 6 Kg.</li> <li>9. Should have Selectable FiO2%: 21% to 100%.</li> <li>10. Product should be European CE certified /USFDA/ BIS Certified.</li> <li>11. Warranty: 3 Years or higher</li> </ol>	42

11	PEEP VALVE, ADJUSTABLE OR SET	<ol style="list-style-type: none"> <li>1. Reusable.</li> <li>2. Designed for use with ventilators and any other systems where PEEP or CPAP is required.</li> <li>3. Transparent to allow visual inspection of performance.</li> <li>4. Monitors Tidal Volume by connecting ventilation meter to outlet port.</li> <li>5. PEEP range from 3 to 20 cm H<sub>2</sub>O pressure.</li> <li>6. Warranty: 1 Year or higher</li> </ol>	42
12	RESUSCITATOR WITH OXYGEN INLET AND MASKS AND AIRWAYS FOR ALL AGES AND OXYGEN RESERVOIR	<ol style="list-style-type: none"> <li>1. Easy Grip manual resuscitator with transparent face-mask;</li> <li>2. Child models (500 to 250ml bag capacity);</li> <li>3. Standard 15-22 mm Swivel connector allows connections to all common masks Endotracheal Tubes;</li> <li>4. Provision to give supplemented oxygen-by-oxygen reservoir providing 100% oxygen;</li> <li>5. Non-re breathing valve enabling the patient to inspire oxygen from the reservoir bag.</li> <li>6. Device should be made of silicon except reservoir</li> <li>7. silicon bellow, non-re - breathing Valve</li> <li>8. Mask (0,1,2 sizes; 3 sets), Oxygen Reservoir bag, 1 meter oxygen tube, Guedel Airway (0,1,2 sizes; 3 sets)</li> <li>9. Bag for storage should be supplied.</li> <li>10. Warranty: 1 Year or higher</li> </ol>	42
13	ELECTRIC PORTABLE SUCTION ASPIRATOR WITH AIR FLOW OF AT LEAST 30 L/MIN AND A VACUUM LEVEL OF AT LEAST 600 MM HG (ISO 10079-1- 1999)	<ol style="list-style-type: none"> <li>1. 0 to 30 Ltrs flow rate Minimum.</li> <li>2. Max. Vacuum / LPM: - 600mm of Hg.</li> <li>3. Portable &amp; with a mounting bracket.</li> <li>4. Collection bottle 1000 ml capacity and sterilizable.</li> <li>5. Overflow protection insured.</li> <li>6. It should be supplied with one set of all standard accessories like Microbial filter, tubing with combination of suction tip &amp; angled connector.</li> <li>7. Product should be CE certified /ISI Marked.</li> <li>8. Warranty: 1 Year or higher.</li> </ol>	42
14	SEAT BELT CUTTER	<ol style="list-style-type: none"> <li>1. 2-in-1 emergency-escape tool provides a window hammer and seat-belt cutter</li> <li>2. Hard tungsten metal double-head hammer for quickly breaking a car-door window</li> <li>3. Sharp blade for easily cutting through a seat belt; plastic handle provides a secure grip</li> <li>4. Protective bracket included for safe storage; store in centre console or driver-side door pocket.</li> <li>5. Warranty: 1 Year or higher</li> </ol>	42
15	IMMOBILIZATION, SET OF FRACTURES	<ol style="list-style-type: none"> <li>1. Immobilization, Set of Fractures.</li> <li>2. Warranty: 1 Year or higher.</li> </ol>	42
16	CERVICAL UPPER SPINAL IMMOBILIZATION DEVICES CERVICAL	<ol style="list-style-type: none"> <li>1. Cervical Upper Spinal Immobilization Devices Cervical.</li> <li>2. Warranty: 1 Year or higher.</li> </ol>	42
17	EXTENDED UPPER SPINAL IMMOBILIZATION EXTRICATION DEVICES OR SHORT SPINAL BOARD (ONE OF THESE)	<ol style="list-style-type: none"> <li>1. Extended Upper Spinal Immobilization Extrication Devices or Short Spinal Board (One of These).</li> <li>2. Warranty: 1 Year or higher</li> </ol>	42

18	Multi-Para Monitor	<ol style="list-style-type: none"> <li>1. Should be Modular Multipara Monitor. Suitable for Neonate, Pediatrics &amp; Adult patients.</li> <li>2. Monitors should have facility to monitor ECG, SpO2, NIBP, Respiration rate, temperature, Heart Rate</li> <li>3. Should have facility to display ECG, RR, SpO2, NIBP, Dual Temperature, Heart Rate</li> <li>4. Should display at least 07 wave form of selected parameters simultaneously.</li> <li>5. Should be able to store &amp; display at least 24 hrs of graphical trends of all parameters.</li> <li>6. The SpO2 technology should be NELLCOR/MASIMO/F AST/SET (BluPro) and sense in low perfusion, hypotension, shivering and motion.</li> <li>7. Monitor must have bright, highly visible minimum 12" or more Colour display with full touch screen facility.</li> <li>8. Should have internal rechargeable battery for 60 minutes or more.</li> <li>9. Monitor should have audio/visible alarm.</li> <li>10. Monitors must have ESU &amp; Defibrillation protection.</li> <li>11. Monitor must be USFDA/EU CE (Notified Body) certified/BIS and Manufacturer should be ISO 13485 Certified. Certificate to be submitted.</li> <li>12. Following accessories should supply with each monitor free of cost: <ol style="list-style-type: none"> <li>A. ECG Cable- 2 Set.</li> <li>B. SPO2 Probe reusable (Adult, Pediatrics, Neonates)- 1 Each</li> <li>C. Temperature Probe -2 Nos. (One core &amp; One peripheral)</li> <li>D. NIBP Cuff (Adult, Adult-Extra-large, Paediatrics, Neonates)- 1 Each</li> <li>E. Wall Mount Stand for each Monitor.</li> <li>F. Moisture Trap with Tubings-20 no. (If Applicable)</li> </ol> </li> <li>13. Warranty-03 Years.</li> </ol>	42
19	Portable Oxygen	Minimum 1 No. of 2.2L Water Capacity Aluminium Cylinder at maximum 150 kgf/cm2 filling pressure manufactured as per IS: BS EN ISO 7866:2012 / ISO 7866:2012 / IS 15660 with Warranty: 1 Year or higher.	42

b) Specification of Medical equipment's to be supplied and installed with ALS Ambulances:

Sl	Description of the item	Specification	Qty
1	OXYGEN CYLINDER (B- TYPE) WITH CYLINDER KEY AND REGULATORS	Minimum 2 Nos. of 10L Water Capacity (B-Type) Cylinders at maximum 150 kgf/cm2 filling pressure manufactured as per IS:7285 along with Cylinder Keys and Regulators with Warranty: 1 Year or higher.	40



2	OXYGEN FLOW METER	<ol style="list-style-type: none"> <li>1. Back Pressure Compensated flow meter should be of accurate gas flow measurement with following feature.</li> <li>2. Control within a range of 0 – 15 LPM. (calibration within <math>\pm 10\%</math>)</li> <li>3. The flow meter body should be made of brass chrome plated materials.</li> <li>4. The flow tube and shroud components should be made of clear, impact resistant polycarbonate.</li> <li>5. The flow tube should have large and expanded 0-5 lpm range for improved readability at low flows</li> <li>6. Warranty: 1 Year or higher</li> </ol>	40
3	HUMIDIFIER	<ol style="list-style-type: none"> <li>1. Humidifier Bottle with 1 Year Warranty.</li> </ol>	40
4	SPINE BOARD STRETCHER	<ol style="list-style-type: none"> <li>1. High Density Poly ethylene – Single piece</li> <li>2. Rigid, Light &amp; Floatable</li> <li>3. Resistant to bumps and corrosion</li> <li>4. Non-absorbent, immune to infiltrations</li> <li>5. Easy to clean – water &amp; soap should be enough.</li> <li>6. X-Ray &amp; MRI compatible</li> <li>7. Net weight approx : 8 Kgs.</li> <li>8. Load bearing minimum: 150 Kg</li> <li>9. Dimensions (LxWxH) approx: 185 x 45 x 5cm tolerance <math>\pm 5\%</math></li> <li>10. Product should be CE certified.</li> <li>11. Warranty: 1 Year or higher.</li> </ol>	40
5	SCOOP STRETCHER	<ol style="list-style-type: none"> <li>1. Should be light, safe and reliable. Made of aluminium alloy</li> <li>2. Clutch Design in the middle so that the stretcher can be divided into left and right halves.</li> <li>3. Adjustable length according to patients' height</li> <li>4. Easy to lock &amp; unlock</li> <li>5. Quick release buckle belts.</li> <li>6. Dimension approx. Size (LxWxH) 225x45x7cm &amp; Min. Size (LxWxH) 168 x 43 x 7cm tolerance <math>\pm 5\%</math></li> <li>7. Net weight approx : 9 Kgs</li> <li>8. Load bearing minimum: 150 Kg</li> <li>9. Product should be CE certified.</li> <li>10. Warranty: 1 Year or higher.</li> </ol>	40

6	Glucometer	<ol style="list-style-type: none"> <li>1. Should have LCD display screen and auto shut off feature when not in use.</li> <li>2. Display of the sugar reading should be in mg/dl.</li> <li>3. Should have reading range/linearity from 10 to 600 mg/d.</li> <li>4. Should have a maximum reading time of less than 60 seconds.</li> <li>5. Should be supplied with 100 nos disposable lancets and 100 strips free of cost.</li> <li>6. Should have the feature of automatic code detection of glucose strips.</li> <li>7. Should have a minimum memory of 100 tests.</li> <li>8. Battery powered:3-volt lithium coin cell battery or 2 x (AAA) Alkaline Batteries.</li> <li>9. Certifications: <ul style="list-style-type: none"> <li>· Should be CDSCO approved.</li> <li>· Should comply with BIS standards.</li> <li>· Should comply with USFDA/European CE /CE-IVD standards incase of non-availability of BIS standards.</li> <li>· Should conform to ISO 13485 quality standards.</li> </ul> </li> <li>10. Glucose Monitoring System should have a haematocrit range of 0-70%.</li> <li>11. The shelf Life of the test strip should be 18 months at the time of supply</li> <li>12. Warranty: 1 Year or higher.</li> </ol>	40
7	AMBU BAG- CHILD (BAG VALUE MASK)	<ol style="list-style-type: none"> <li>1. Made of High-grade Silicon Rubber, Reusable &amp; Autoclavable at 134°C supplied with Face Mask Size-Child size &amp; Reservoir Bag, Oxygen Tube.</li> <li>2. Complete Packed in P.V.C. Carry Bag with zip Lock.</li> <li>3. Should have ISO 13485 certificates.</li> <li>4. Warranty: 1 Year or higher.</li> </ol>	40
8	AMBU BAG- ADULT (BAG VALUE MASK)	<ol style="list-style-type: none"> <li>1. Made of High-grade Silicon Rubber, Reusable &amp; Autoclavable at 134°C supplied with Face Mask Size-Adult size &amp; Reservoir Bag, Oxygen Tube.</li> <li>2. Complete Packed in P.V.C. Carry Bag with zip Lock.</li> <li>3. Should have ISO 13485 certificates.</li> <li>4. Warranty: 1 Year or higher.</li> </ol>	40
9	ARTERY FORCEPS 6"	<ol style="list-style-type: none"> <li>1. Size:6"(Straight)</li> <li>2. High-quality stainless-steel corrosion resistant, long durability and reusable.</li> <li>3. Warranty: 1 Year or higher.</li> </ol>	40
10	FORCEPS PLAIN 6"	<ol style="list-style-type: none"> <li>1. Size:6"(Plain)</li> <li>2. High-quality stainless-steel corrosion resistant, long durability and reusable.</li> <li>3. Warranty: 1 Year or higher.</li> </ol>	40

11	BP APPRATUS	<ol style="list-style-type: none"> <li>Should be able to measure blood pressure in adult as well as paediatric patients.</li> <li>Pressure measurement range should be 0 to 300 mm Hg systolic and 40 to 200 mm diastolic.</li> <li>Pressure measurement accuracy of +/- 3 to 5 mm Hg</li> <li>Should be supplied with standard cuff as below: Paediatric, Infant, Adult size blood pressure cuffs.</li> <li>Warranty: 1 Year or higher.</li> </ol>	40
12	NEBULISOR MACHINE	<ol style="list-style-type: none"> <li>Motion Tolerant and for continuous use in Pre-Hospital transportation.</li> <li>Maximum pressure 35 psi.</li> <li>Air power: 14 liters per minute</li> <li>Aerosol output: 106 per minute</li> <li>Residual volume: 1.24 ml</li> <li>Droplet size: MMAD 3.3 microns</li> <li>Filling volume: maximum 7ml</li> <li>Noise level: 55 db.</li> <li>Provisions for fixing/Hanging in the Ambulance.</li> <li>Check valve to protect the device against contamination due to backward inhalation.</li> <li>Unbreakable lids.</li> <li>Dust filter.</li> <li>Product should be CE certified.</li> <li>Warranty: 1 Year or higher.</li> </ol>	40
13	NEEDLE AND SYRINGE DESTROYER	<ol style="list-style-type: none"> <li>Should have Stainless Steel Body of SS-304 grade/ABS.</li> <li>Should have minimum 400 ml Container</li> <li>Should have stainless steel alloy adjustable blade (Non-Magnetic)</li> <li>Should have Easy removable cap</li> <li>Should have plastic holder at the end of the cutter</li> <li>Needle Container should have Bio-Hazard symbol of (Black/Red) on the front &amp; Back.</li> <li>Manufacturer should be ISO 13485 certified and Certificate need to be submitted.</li> <li>Warranty: 1 Year or higher.</li> </ol>	40
14	PULSE OXYMETER	<ol style="list-style-type: none"> <li>Pulse Oximeter: finger tip</li> <li>Warranty: 1 Year or higher.</li> </ol>	40
15	SCISSOR STREIGHT	<ol style="list-style-type: none"> <li>Size:6"</li> <li>High-quality stainless-steel corrosion resistant, long durability and reusable.</li> <li>Warranty: 1 Year or higher.</li> </ol>	40
16	SCISSORS 6" WITH ROUND TIP	<ol style="list-style-type: none"> <li>Size:6"</li> <li>High-quality stainless-steel corrosion resistant, long durability and reusable.</li> <li>Warranty: 1 Year or higher.</li> </ol>	40

17	STETHOSCOPE	<ol style="list-style-type: none"> <li>1. Chest Piece Stainless Steel Dual Head Diaphragm</li> <li>2. Sensitive Diaphragm, Non-Chilling Snap-on Ring Combination of 2 in 1 Pediatric and Neo Natal.</li> <li>3. 2-way Tube with extra Gloss Finish, extra Soft Sealing Ear Knobs.</li> <li>4. 6 mm Stainless Steel comfortable head frame.</li> <li>5. Should have ISO 9001 Certified.</li> <li>6. Warranty: 1 Year or higher.</li> </ol>	40
18	SUCTION PUMP HAND OPERATED	<ol style="list-style-type: none"> <li>1. Bottle with 200 ml capacity with 1 Year Warranty.</li> </ol>	40
19	THERMOMETER DIGITAL	<ol style="list-style-type: none"> <li>1. Temperature display: - 3-digit, +°C (°F) display in 0.1 degree increment.</li> <li>2. Measurement accuracy: - <math>\pm 0.1^{\circ}\text{C}</math> (32-42°C); <math>\pm 2^{\circ}\text{F}</math> (89.6-107.6°F).</li> <li>3. Should beep when measurement gets completed.</li> <li>4. Should be supplied with protective case, test battery.</li> <li>5. The manufacturer should have valid ISO 13485 Certificate.</li> <li>6. Warranty: 1 Year or higher.</li> </ol>	40
20	TONGUE DEPRESSOR WOODEN	<ol style="list-style-type: none"> <li>1. Tongue depressor should be made of a high-quality wood.</li> <li>2. 6" size to use on children, adults, and seniors.</li> <li>3. Warranty: 1 Year or higher.</li> </ol>	40
21	TOOTHED FORCEPS 6"	<ol style="list-style-type: none"> <li>1. Size: 6" (Straight)</li> <li>2. High-quality stainless-steel corrosion resistant, long durability and reusable.</li> <li>3. Warranty: 1 Year or higher.</li> </ol>	40
22	Portable Oxygen	Minimum 1 No. of 2.2L Water Capacity Aluminium Cylinder at maximum 150 kgf/cm <sup>2</sup> filling pressure manufactured as per IS: BS EN ISO 7866:2012 / ISO 7866:2012 / IS 15660 with Warranty: 1 Year or higher.	40
23	Electric Portable Suction Aspirator	<ol style="list-style-type: none"> <li>1. 0 to 30 Ltrs flow rate Minimum.</li> <li>2. Max. Vacuum / LPM: - 600mm of Hg.</li> <li>3. Portable &amp; with a mounting bracket.</li> <li>4. Collection bottle 1000 ml capacity and sterilizable.</li> <li>5. Overflow protection insured.</li> <li>6. It should be supplied with one set of all standard accessories like Microbial filter, tubing with combination of suction tip &amp; angled connector.</li> <li>7. Product should be CE certified / ISI Marked.</li> <li>8. Warranty: 1 Year or higher.</li> </ol>	40

c) Following item category will be considered as same or similar Category Products for this bid.

- Ambulances without Equipment
- Customised Ambulance with Equipment

d) Bid Specific Authorization Letter from the OEM (Ambulance and each equipment's) signed by authorized signatory heaving Power of Attorney/ Board Resolution mentioning the Bid reference no is to be submitted by the bidder if the bidder is not an OEM (PoA/ Board Resolution also need to be submitted).

e) The rate of the Ambulances should quote along with One (1) Year Insurance with Zero depreciation, R

egistration and Branding.

- d) Registration of the Ambulances will be in the name of Mission Director, National Health Mission, Assam.
- e) The registration and installation of name plate of the Ambulances should be done by the selected bidder.
- f) Parking location for the Ambulances will be fixed by National Health Mission, Assam at the time of delivery. Intimation should be done before 7 days of delivery of Ambulances.
- g) Timely Servicing / rectification of defects during warranty period should be done within 7 days excluding major accidental repair.
- h) 75% districts of Assam should have operational Service Centre for easy and effective maintenance of Ambulances. List of operational Service Centre need to be submitted.
- i) Format for submission of Technical Compliance of specification (Ambulance and each equipment's):

(In OEMs letter Head)				
Technical Compliance of Specification				
Name of the Item (Ambulance/ Equipment)				
Make & Model of offered Product		Make: Model:		
Sl	Specification mentioned in the Bid Document	Specification of the offered Product	Attested Certificates (If mentioned in Bid Document)	Compliance (Yes/ No)
1				
2				
3				
 Signature : Name & Designation : Date : Name & Address of the Firm : Email Id : Contact No :				

#### 24. Sample Clause

After award of contract – Successful Bidder shall have to get advance sample approved from buyer before bulk manufacturing / starting bulk supplies. Successful Bidder shall submit

1

samples for Buyer's approval, within 21 days of award of contract. Buyer shall, as per contract specifications framework, either approve the advance sample or will provide complete list of modification required in the sample within 7 days of receipt of advance sample. Seller shall be required to ensure supply as per approved sample with modifications as communicated by Buyer. If there is delay from buyer side in approval of advance sample – the delivery period shall be refixed without LD for the period of delay in sample approval. In case, the sample is found to have major deviations / not conforming to the Contract specifications, the buyer at its discretion may call for fresh samples for approval before allowing bulk supplies or may terminate the contract after notifying the deviations to the seller.

Unless otherwise provided in the contract, all samples required for test shall be supplied by the contractor

free of cost. Where under the contract, the contractor is required to submit an advance sample, any expenses incurred by the contractor on or in connection with the production of stores in bulk, before the sample has been approved unconditionally shall be borne by the Seller and he shall not claim any compensation in the event of such sample being found unacceptable by the Buyer / Consignee.

## अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**